



Missouri Department of Health and Senior Services

P.O. Box 570, Jefferson City, MO 65102-0570 Phone: 573-751-6400 FAX: 573-751-6010  
RELAY MISSOURI for Hearing and Speech Impaired 1-800-735-2966 VOICE 1-800-735-2466

Margaret T. Donnelly  
Director



Jeremiah W. (Jay) Nixon  
Governor

Missouri Department of Health and Senior Services Breath Alcohol Program

***CERTIFICATE OF CALIBRATION***

This report shall document that the instrument described herein was examined and tested against NIST-traceable reference standards in accordance with Missouri Department of Health and Senior Services State Public Health Laboratory Breath Alcohol Program's Testing Requirements for the Recalibration of Simulators.

AGENCY: Sikeston Department of Public Safety

ADDRESS: 215 N. New Madrid

CITY: Sikeston, MO

ZIP: 63801

**SIMULATOR INFORMATION**

Serial Number: SD 2245

Manufacturer: Guth

Model Number: 10-4D

**CALIBRATION RESULTS**

<u>Reference Temperature</u>	<u>Simulator Temperature</u>	<u>Tolerance</u>	<u>Bias</u>	<u>In Tolerance</u>	<u>Combined Uncertainty</u>
33.98	34.0	±0.1 °C	-0.02 °C	YES	±0.04 °C

This calibration was performed with NIST-Traceable Thermometer SN: 304447

This calibration was performed by: Brian Lutmer

This calibration was performed: 03/12/2013